FREND[™] System

Quick manual

Code chip installation

1 Turn on the power switch.

2 Check the lot number.






- Insert the cartridge code chip into the code chip slot in the direction of the arrow.
- 4 Press the **SETUP** button on the main screen.





- 5 Press the **Code Chip** button on the setup screen.
- The cartridge code chip installation is complete. Press the **OK** button.





FREND[™] System



Running a Test

- 1 Press the **TEST** button on the main screen.
- Type the patient ID and press the **Enter** button.






- Insert the cartridge into the cartridge slot.
- Wait until the test is done automatically. After the reaction is complete, the FREND™

 System reads the value.





When the test is done, the cartridge comes out and the results are displayed automatically.



A Caution

- Each cartridge box contains one cartridge code chip. The code chip works for all cartridge in the same box.
- If you want to remove the cartridge code chip from the FREND™ System, press the cartridge code chip gently.
- Allow the cartridge at room temperature for 15 ~ 30 minutes before use.
- Avoid air bubbles. Contamination, hemolysis, and undesirable amount of sample can lead to inaccurate results.
- Check the direction of the cartridge before insertion. And also check if the insertion is complete. (Don't' insert upside down)